



### Seraphin Test Measure

A Division of Pemberton Fabricators, Inc.

PO Box 227, 30 Indel Ave.

Rancocas, NJ 08073

**Phone:** 609-267-0922 • **Fax:** 609-261-2546

**Email:** [sales@seraphinusa.com](mailto:sales@seraphinusa.com) • **Website:** [www.seraphinusa.com](http://www.seraphinusa.com)

## Item # EESS0020LB, 20 LITER TEST MEASURE

20 LITER TEST MEASURE



List Price

**\$917.00**

### Specifications

#### CALIBRATION TYPE

Factory CALIBRATION by standards traceable to National Institute of Standards and Technology

#### CONSTRUCTION METHOD

Built to OIML R-120, N.I.S.T. Handbook 105-3 Standards and API Manual of Petroleum Measurement Standards Chapter 4

#### DESIGN

Reinforcing bands to prevent distortion of prover when full with liquid and when being transported.

#### GAUGE ASSEMBLY

Equipped with a (liquid -level) gauge tube mounted on the side of the neck

#### GRIP

Rubber grip on test measure handle

#### MATERIAL

TYPE 304 STAINLESS STEEL

#### NOMINAL CAPACITY

20 Liters - calibrated "to deliver" at 15 Deg C

#### NOMINAL DIAMETER

11 inches

#### NOMINAL HEIGHT

26 inches

<b>REQUIRED SCALE RANGE</b>	± 200 milliliters. The gauge scale shall be graduated both above and below the nominal capacity graduation by an amount not less than 1.5 % of the test measure
<b>SCALE RESOLUTION</b>	10 MILLILITERS
<b>WEIGHT (DRY)</b>	15 LBS.
<b>TOLERANCE</b>	0.02% Scale Accuracy
<b>VOLUME INDICATION</b>	ALUMINUM GRADUATED SCALE
<b>BOTTOM</b>	Concave and of adequate thickness to prevent distortion when filled.